Mettler-Toledo Report Address Service LabTec Im Langacher 44 8606 Greifensee N/A

METTLER TOLEDO

Certificate Japanese Pharmacopoeia General Tests 9.62 Measuring Instruments, Appliances, "Balances and Weights"

Custon	ner					
Cor	mpany:	Sample Custor	er			
Ado	dress:	Sample Addres	6			
City	<i>/</i> :	Sample City			Contact:	N/A
Zip/	/Postal:	Sample			Order Number:	Costumer order no.
Sta	te/Province:	Sample State				
Weighi	ng Device					
Mai	nufacturer:	Mettler Toledo			Instrument Type:	Weighing Instrument
Мо	del:	XPR225DR			Asset Number:	N/A
Ser	ial No.:	Sample SNR			Alternate Asset No.:	N/A
Bui	lding:	Sample Building			Terminal Model:	PRAT
Flo	or:	Floor no.			Terminal Serial No .:	Sample SNR
Roc	om:	Room no.			Terminal Asset No.:	N/A
	Range	Max. Capacity 🔨	F	Readabilit	y (d)	
	1	121 g		0.00001	g	
	2	220 g		0.0001	g	
Proced	lure					
Ref	erence Document:	\sim	Japanese P Scales (Bal	harmacopo ances) and	peia General Test Methods I Weights	9.6 Weighing Instruments/Equipment
ME	METTLER TOLEDO Work Instruction:			eia Certifica	ate WI 10000027820	
This	s certificate contain	s measurements for	As Found and A	s Left tests	i.	
As	As Found Test Date:27-As Left Test Date:27-Issue Date:27-		27-Mar-2025 27-Mar-2025 27-Mar-2025		Service Technician:	111
As						Ambros Kohler
Issu						
Nex	t Test Date:	3	1-Mar-2026			

Summary of Results

	Repeatability	As Found	As Left	
Test	Smallest Net Weight	Tare Load	Assessment	Assessment
RP_SNW_0.05000g	0.05000 g	N/A	✓	✓
	Accuracy			As Left
	✓	✓		

Measurement Results

Repeatability

NW_0.05000g				
Smallest Net Weight:	0.05000 g	Tare Vessel ID:	N/A	
Test Load:	10 g	Tare Vessel Descri	iption: N/A	
Tare Load:	N/A			
	As Found	As Left	• As Found	
1	10.00001 g	9.99999 g	◆ As Left >5d	
2	10.00001 g	9.99999 g	10 4d 2	
3	10.00001 g	9.99999 g	30	
4	10.00000 g	10.00000 g		
5	10.00000 g	10.00000 g		
6	10.00000 g	10.00000 g		
7	10.00000 g	10.00000 g	000	
8	10.00000 g	10.00000 g	8	
9	10.00000 g	10.00000 g		
10	10.00001 g	10.00000 g	7 5	
Mean Value	10.000004 g	9.999997 g	6	
Standard Deviation	0.000005 g	0.000005 g	The "d" in the graph represents the readability of	
Calculation ¹	0.0207 %	0.0193 %	the range/interval in which the test was performed	d.
Assessment ²	0.02 %	0.02 % 🗸	The results of this graph are based upon the	
Requirement	0.10 %	0.10 %	absolute values of the differences from the mean	
Minimum Weight ³	0.01033 g	0.00966 g	value.	

¹The following value is calculated: 2 * standard deviation / smallest net weight. If the standard deviation s is smaller than the rounding error of 0.41*d where d is the readability of the range/interval in which the test was performed, then s is replaced by 0.41*d.

² The assessment is carried out after the calculated value is mathematically rounded to the readability of the requirement of 0.10 %.

³ Minimum weight = 2000 * s. If the calculated standard deviation s is smaller than the rounding error of 0.41*d where d is the readability of the range/interval in which the test was performed, then s is replaced by 0.41*d. In this case, minimum weight = 2000 * 0.41 * d.

All intermediate calculations are performed in the software to 16 decimal places.

Accuracy

Sensitivity

	As Found	As Left	
Test Load	200 g	200 g	
CMV	200.0000 g	200.0000 g 200.0001 g	
Indication	199.9999 g		
Doviation 1	-0 0001 a	0.0001 g	
Deviation	-0.000 i g 🗸	0.000 r g 🗸 🗸	
Requirement	0.1000 g	0.1000 g	

¹ The sensitivity test is passed if the absolute value of the deviation ≤ 0.05 % of the test load value. The requirement for the assessment of sensitivity is 0.05 %. This ensures adherence to the overall accuracy requirement of 0.10 % because other balance properties might also limit the accuracy of the instrument.

Reference Weights

All weights used for metrological testing are traceable to national or international standards. The weights were calibrated and certified by an accredited calibration laboratory.

Weight Set 1: OIML E₂

Weight Set No.:	E2 2013	Date of Issue:	24-Jul-2013	
Certificate Number:	1234567	Calibration Due Date:	21-Jun-2025	
una a ul ca				

Remarks

The user is responsible for maintaining the configuration (settings) of the balance which was used when the assessment was performed.

This document is issued to record completion of the work performed by METTLER TOLEDO on the subject device in accordance with agreed standards. It does not guarantee the continued performance of the subject device. Any measurements recorded are based on the subject device's performance at a given time as tested by METTLER TOLEDO and, except where explicitly stated otherwise, do not express an opinion as to the sufficiency of any customer designed procedures used to test the device. This document is not a warranty, either implied or express. METTLER TOLEDO expressly disclaims any liability arising from the use of the information in this document for any purpose other than as specified herein.